

primary studies - published RCT

# Influence of the Vibralung Acoustical Percussor on pulmonary function and sputum expectoration in individuals with cystic fibrosis.

Code: PM29697011 Year: 2018 Date: 2018

Author: Wheatley CM

## Study design (if review, criteria of inclusion for studies)

2 randomized srudies

## **Participants**

Over two separate studies, 10 and 11 mild to moderate CF patients were recruited for study I and II, respectively.

#### Interventions

Study I: Vibralung was used for 20 min with either no sound (NS: PEP only) or sound (S: PEP and sound) on randomized visits. Study II: over 5 days of in-hospital therapy, the Vibralung or vibratory vest therapy (Vest) were used for two therapy sessions per day, with sputum collected for 20 min following each therapy and pulmonary function accessed pre and post each 5-day period (days 1-5 or 7-11) in a randomized crossover design.

## **Outcome measures**

Study I: Pulmonary function, diffusion capacity of the lungs for carbon monoxide and nitric oxide (DLCO/DLNO), symptoms, and peripheral oxygen saturation (SpO2) were measured at baseline and at 1 and 4 h post treatment. Expectorated sputum was collected over 4 h post treatment. Study II: sputum collected for 20 min following each therapy and pulmonary function accessed pre and post each 5-day period (days 1-5 or 7-11) in a randomized crossover design.

#### Main results

Vibralung usage resulted in no change from baseline to 4 h post in pulmonary function, SpO2 or symptoms (p > 0.05). At 4 h post therapy, the DLCO- and DLNO-derived measure of alveolar-capillary unit function (DM/ VC) showed improvement (DM/ VC = 12.5 +/- 5.5 versus 7.3 +/- 18.8% change, S versus NS) with no difference between S and NS (p = 0.74). Sputum expectoration was similar between S and NS conditions (wet sputum = 10.5 +/- 4.6 versus 9.9 +/- 3.2 g, S versus NS, p = 0.25). There were no differences in the improvement in pulmonary function between Vibralung and Vest during either 5-day period during the hospital stay.

# Authors' conclusions

Vibralung was well tolerated and caused no detrimental changes in pulmonary function metrics. The Vibralung appears to be a safe ACT in individuals with CF.

http://dx.doi.org/10.1177/1753466618770997

#### See also

Ther Adv Respir Dis. 2018 Jan-Dec;12:1753466618770997. doi: 10.1177/1753466618770997.

# Keywords

Airway clearance technique; exercise; non pharmacological intervention - devices OR physiotherapy; Percussion; Vibration; Inhalation OR nebulised; Postural Drainage; Chest physiotherapy; oscillating devices;